

CM 22437 - STERILE SALINE 0.90%

INTENDED USE

Used as a diluent

PRODUCT SUMMARY AND EXPLANATION

Sterile Saline 0.90% is an isotonic diluent used for dilution of bacterial cells to provide a concentration suitable for microscopic observation, determination of cell numbers, analysis for genetic or metabolic properties, washing cells preparatory to study, or preparation of standardized inoculums.

COMPOSITION

| Ingredients | Gms / Ltr |
|-----------------|-----------|
| Sodium chloride | 9.0 |

PRINCIPLE

Sterile saline, 0.90% contains sodium chloride which helps to maintain cell integrity and viability in a bacterial cell suspension.

INSTRUCTION FOR USE

Inoculate the sample and incubate at specified temperature and time.

QUALITY CONTROL SPECIFICATIONS

| | | |
|-------------------------------|---|---------------------------|
| Appearance of prepared medium | : | Colourless solution |
| Quantity of Medium | : | 10 ml of medium in tubes. |
| pH (at 25°C) | : | 7.00 |
| Sterility Check | : | Passes release criteria |

INTERPRETATION

Cultural characteristics were observed after incubation.

| Microorganism | ATCC | Inoculum (CFU/ml) | Growth | Incubation Temperature | Incubation Period |
|-------------------------------|-------|-------------------|-----------|------------------------|-------------------|
| <i>Escherichia coli</i> | 25922 | 50-100 | Luxuriant | 35-37°C | 18-24 hours |
| <i>Staphylococcus aureus</i> | 25923 | 50-100 | Luxuriant | 35-37°C | 18-24 hours |
| <i>Pseudomonas aeruginosa</i> | 27853 | 50-100 | Luxuriant | 35-37°C | 18-24 hours |

PACKAGING:

Pack of 50 Ready-To-Use Liquid Medium tubes containing 9ml and 10 ml in each tube.

Pack of 25 Ready-To-Use Liquid Medium tubes containing 10 ml in each tube.

STORAGE

On receipt, store tubes in the dark at 10-25 °C. Avoid freezing and overheating. Do not open until ready to use. Minimize exposure to light. Tubed media stored as labeled until just prior to use may be inoculated up to the expiration date and incubated for the recommended incubation times. Allow the medium to warm to room temperature before inoculation.



DISPOSAL

User must ensure safe disposal by autoclaving and/or incineration of used or unusable preparations of this product. Follow established laboratory procedures in disposing of infectious materials and material that comes into contact with clinical sample must be decontaminated and disposed of in accordance with current laboratory techniques.

REFERENCES

1. Koch, A.L. (with P. Gerhardt). 1994. 11.3.1, Diluents, p. 255. In P. Gerhardt (ed.), Methods for general and molecular bacteriology. American Society for Microbiology, Washington, D.C.
2. Clinical and Laboratory Standards Institute. 2006. Approved Standard: M2-A9. Performance standards for antimicrobial disk susceptibility tests. 9th ed. CLSI, Wayne, Pa.
3. Clinical and Laboratory Standards Institute. 2006. Approved Standard: M7-A7. Methods for dilution antimicrobial susceptibility tests for bacteria that grow aerobically, 7th ed. CLSI, Wayne, Pa.
4. Thomson, Jr., R.B., and J.M. Miller. 2003. Specimen collection, transport, and processing: bacteriology, p. 286–330. In P.R. Murray, E.J. Baron, J.H. Jorgensen, M.A. Tenover, and R.H. Tenover (ed.), Manual of clinical microbiology, 8th ed., American Society for Microbiology, Washington, D.C.



Quantity



Lot / Batch Number



Temperature Unit



Manufacturer



Best Before



Certification of
Good Manufacturing Practices



Catalogue No.



Authorized Representative



European Conformity



Consults Instructions for use :



QR
Code



For In Vitro Diagnostic Use

NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

***ForLabUse Only**