

CM 22553 - MANNITOL SALT AGAR PLATE

INTENDED USE

For selective isolation of pathogenic Staphylococci.

PRODUCT SUMMARY AND EXPLANATION

MANNITOL SALT AGAR is used for the selective isolation and enumeration of Staphylococci species. Chapman formulated Mannitol Salt Agar to isolate staphylococci by inhibiting growth of most other bacteria with a high salt concentration. Pathogenic staphylococci, i.e. *Staphylococcus aureus* (coagulase test positive) is able to ferment Mannitol but other coagulase negative *Staphylococcus* are not. So, if that particular specimen contains *S. aureus*, it ferments mannitol and changes the pH of medium to acidic. As MSA contains phenol red as a pH indicator, at pH levels below 6.9, the medium is a yellow colour. But if Coagulase negative staphylococci grow, they can't ferment mannitol, so the colour of the media around the bacterial colony does not change to yellow, it appears pink.

Note: *Staphylococcus saprophyticus* (coagulase negative *Staphylococcus*) may ferment mannitol, producing yellow halo around colonies in MSA thus resembling with *S. aureus*

COMPOSITION

Ingredients	Gms / Ltr
Sodium chloride	75.000
Agar	15.000
Mannitol	10.000
Peptic digest of animal tissue	5.000
Pancreatic digest of casein	5.000
Beef extract	1.000
Phenol red	0.025

PRINCIPLE

The medium contains Peptic digest of animal tissue, Pancreatic digest of casein and Beef extract which are the source of amino acid and nitrogen. Mannitol is the carbohydrate energy source and Phenol red is the pH indicator. The 7.5% concentration of sodium chloride results in the partial or complete inhibition of bacterial organisms other than Staphylococci. Agar is a solidifying agent. The dye turns the medium colour yellow when high acidic condition prevails in the medium due to mannitol fermentation. After the recommended incubation period, the plates should show isolated colonies in streaked areas and confluent growth in areas of heavy inoculation. Coagulase - positive Staphylococci produce growth of yellow colonies with yellow zones. Coagulase negative Staphylococci produce small red colonies with no color change to the medium.

INSTRUCTION FOR USE

Either streak, inoculate or surface spread the test inoculum aseptically on the plate.

QUALITY CONTROL SPECIFICATIONS

Appearance	:	Red coloured Medium
Quantity of Medium	:	25ml of medium in 90mm plates.
pH (at 25°C)	:	7.4 ± 0.2
Sterility Check	:	Passes release criteria



INTERPRETATION

Cultural response was observed after incubation.

Microorganism	ATCC	Inoculum (CFU/ml)	Growth	Recovery	Appearance of colony	Incubation Temperature	Incubation Period
<i>Staphylococcus aureus</i>	6538	50-100	Luxuriant	>=70%	yellow/white colonies surrounded by yellow zone	30-35°C	18-72 hours
<i>Staphylococcus epidermis</i>	12228	50-100	Luxuriant	>=70%	red colour	30-35°C	18-72 hours
<i>Escherichia coli</i>	25922	≥ 1000	Inhibited	-	---	30-35°C	18-72 hours
* <i>Klebsiella aerogenes</i>	13048	≥ 1000	Inhibited	-	---	30-35°C	18-72 hours

* Formerly known as *Enterobacter aerogenes*

PACKAGING:

Double layered packing containing 5 No. of plates with one silica gel desiccant bag packed inside it.

STORAGE

On receipt, store the plates at 15–30 °C. Avoid freezing and overheating. Do not open until ready to use. Prepared plates stored in their original sleeve wrapping until just prior to use may be inoculated up to the expiration date and incubated for recommended incubation times. Allow the medium to warm to room temperature before inoculation.













Product Deterioration: Do not use plates if they show evidence of microbial contamination, discoloration, drying, cracking or other signs of deterioration.

DISPOSAL

After use, prepared plates, specimen/sample containers and other contaminated materials must be sterilized before discarding.

REFERENCES

1. Chapman G.H., 1945, J. Bact., 50:201.
2. Marshall R. (Ed.), 1992, Standard Methods for the Examination of Dairy Products, 16th ed., APHA, Washington, D.C.
3. Bacteriological Analytical Manual, 1995, Food and Drug Administration, 8th ed., AOAC, International, U.S.A

 Quantity	 Lot / Batch Number	 Temperature Unit	 Manufacturer	 Best Before	 Certification of Good Manufacturing Practices
 Catalogue No.	 Authorized Representative <small>MedNet GmbH Birkensweg 10, 48163 Münster, Germany</small>	 European Conformity	 QR Code	 Consults Instructions for use :	 For In Vitro Diagnostic Use

NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

