

CM 20355 – C.L.E.D. AGAR BASE W/O INDICATOR

INTENDED USE

For isolation, enumeration and presumptive identification of bacterial flora in the urinary tract.

PRODUCT SUMMARY AND EXPLANATION

Onasolid medium, Sandy's reported that swarming of Proteus species could be controlled by restricting the electrolytes. Formerly swarming of Proteus was controlled by adding alcohol, surface-active agent, sodium azide, boric acid etc. to the medium. Later on Sandys medium was modified by Mackey and Sandy's, by replacing mannitol by lactose and sucrose and elevating concentration of agar and bromothymol blue. This formulation was further modified by the same authors and called C.L.E.D. (Cystine-Lactose-Electrolyte-Deficient) by deleting the sucrose and by including L-cystine for promoting the growth of cystine dependent dwarf coliform colony. This medium is recommended for use in urine bacteriology, promoting the growth of all urinary pathogens. C.L.E.D. Medium is also recommended for dipstick procedures and as dip inoculum transport medium for urine specimens.

Bacteriuria may be quantitated by inoculating the surface of an agar medium by proper dilution. Inoculate the medium immediately after urine collection. It can also be inoculated by calibrated loop or duplicate dilution pour plate methods. Shigella species may not grow on this medium. Initiation of antibiotic therapy, before collection sample, low urine pH (less than 5) etc. may result in low bacterial count from infected patients.

COMPOSITION

Ingredients	Gms / Ltr
Peptone	4.000
Tryptone	4.000
Beef extract	3.000
Lactose	10.000
L-Cystine	0.128
Agar	15.000

PRINCIPLE

Peptone, beef extract and tryptone provides nitrogen and carbon source, long chain amino acids, vitamins and other essential growth nutrients. Lactose is the fermentable sugar. L-cystine supports the growth of dwarf coliform colony. Bromo thymol blue is the pH indicator which turns yellow at acidic pH.

INSTRUCTION FOR USE

- Dissolve 36.1grams in 998 ml purified/ distilled water.
- Add rehydrated contents of 1 vial of Bromo Thymol Blue Supplement.
- Heat, to boiling, to dissolve the medium completely.
- Sterilize by autoclaving at 15 psi pressure (121° C) for 15 minutes.
- Mix well and pour into sterile Petri plates.

QUALITY CONTROL SPECIFICATIONS



Appearance of Powder : Cream to yellow homogeneous free flowing powder.
 Appearance of prepared medium : With addition of Bromo Thymol Blue Supplement : Green coloured clear to slightly opalescent gel forms in Petri plates.
 pH (at 25°C) : 7.3±0.2

INTERPRETATION

Cultural characteristics observed after incubation.

Microorganism	ATCC	Inoculum (CFU/ml)	Growth	Recovery	Colour of colony	Incubation Temperature	Incubation Period
Escherichia coli	25922	50-100	Good-luxuriant	>=50%	Yellow, opaque, center slightly deeper yellow	35-37°C	18-24 Hours
Enterococcus faecalis	29212	50-100	Good-luxuriant	>=50%	Slight yellowish or greenish	35-37°C	18-24 Hours
Klebsiella pneumoniae	13883	50-100	Good-luxuriant	>=50%	Yellow to whitish blue	35-37°C	18-24 Hours
Proteus vulgaris	13315	50-100	Good-luxuriant	>=50%	Blue	35-37°C	18-24 Hours
Staphylococcus aureus subsp. aureus	25923	50-100	Good-luxuriant	>=50%	Deep yellow	35-37°C	18-24 Hours
Salmonella Typhi	6539	50-100	Good-luxuriant	>=50%	Bluish	35-37°C	18-24 Hours

PACKAGING:

Inpacksize of 100 gm and 500 gm bottles.

STORAGE

Dehydrated powder, hygroscopic in nature, store in a dry place, in tightly-sealed containers between 25-30°C and protect from direct sunlight. Under optimal conditions, the medium has a shelf life of 4 years. When the container is opened for the first time, note the time and date on the label space provided on the container. After the desired amount of medium has been taken out replace the cap tightly to protect from hydration.

Product Deterioration: Do not use if they show evidence of microbial contamination, discoloration, drying or any other signs of deterioration.

DISPOSAL













After use, prepared plates, specimen/sample containers and other contaminated materials must be sterilized before discarding.

REFERENCES

1. Benner E. J., 1970, Appl. Microbiol., 19(3), 409.
2. Dixon J. M. S. and Clark M. A., 1968, Conc. Med. Assoc. J., 99 (15).
3. MacFaddin J. F., 1985, Media for Isolation-Cultivation-Identification-Maintenance of Medical Bacteria, Vol. I, Williams and Wilkins, Baltimore.
4. Mackey and Sandys, 1965, Br. Med. J., 2:1286.



5. Mackey and Sandys, 1966, Br. Med. J., 1:1173.
6. Sandys, 1960, J. Med. Lab. Technol., 17:224.

 GMP Good Manufacturing Practices Certified	 IVD For In Vitro Diagnostic Use	 QTY. Quantity	 LOT/ B. NO. Lot / Batch Number	 REF Catalogue Number	 Manufacturer
 Temperature Unit	 EC REP Authorized Representative <small>MedNet GmbH Barkstrasse 10, 48163 Maastricht, Germany</small>	 European Conformity	 QR Code	 Consults Instructions for Use	 Best Before

NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

*For LabUse Only

