

CM 21,318 – NUTRIENT BROTH

INTENDED USE

For general cultivation of less fastidious microorganisms, can be enriched with blood.

PRODUCT SUMMARY AND EXPLANATION

Nutrient media are basic culture media used for maintaining microorganisms, cultivating fastidious organisms by enriching with serum or blood and are also used for purity checking prior to biochemical or serological testing. Nutrient Broth has the formula originally designed for use in the Standard Method for Examination of Water and Waste water. It is one of the several non-selective media useful in routine cultivation of microorganisms. It can be used for the cultivation and enumeration of bacteria which are not particularly fastidious. Addition of different biological fluids such as horse or sheep blood, serum, egg yolk etc. makes it suitable for the cultivation of related fastidious organisms.

COMPOSITION

Ingredients	Gms / Ltr
Peptone	5.000
Beef extract	1.500
Yeast extract	1.500
Sodium chloride	5.000

PRINCIPLE

The medium consists of Peptone, Beef extract and yeast extract that provide the necessary nitrogen compounds, carbon, vitamins and also some trace ingredients necessary for the growth of bacteria. Sodium chloride maintains the osmotic equilibrium of the medium.

INSTRUCTION FOR USE

- Dissolve 13.0 grams in 1000 ml purified/distilled water.
- Heat, if necessary, to dissolve the medium completely.
- Dispense into tubes or flasks as desired.
- Sterilize by autoclaving at 15 psi pressure (121°C) for 15 minutes.

QUALITY CONTROL SPECIFICATIONS

Appearance of Powder	: Cream to yellow homogeneous free flowing powder.
Appearance of prepared medium	: Light yellow coloured clear to slightly opalescent solution.
pH (at 25°C)	: 7.4 ± 0.2

INTERPRETATION

Cultural characteristics observed after incubation.

Microorganism	ATCC	Inoculum (CFU/ml)	Growth	Incubation Temperature	Incubation Period
---------------	------	-------------------	--------	------------------------	-------------------



Escherichia coli	25922	50-100	Good-luxuriant	35-37°C	18-48 Hours
Salmonella Typhi	6539	50-100	Good-luxuriant	35-37°C	18-48 Hours
Staphylococcus aureus aubs.p.aureus	25923	50-100	Good-luxuriant	35-37°C	18-48 Hours
Streptococcus pyogenes	19615	50-100	Good-luxuriant	35-37°C	18-48 Hours
Pseudomonas aeruginosa	27853	50-100	Good-luxuriant	35-37°C	18-48 Hours

PACKAGING:

Inpacksizeof100 gm and 500 gm bottles.

STORAGE

Dehydrated powder, hygroscopic in nature, store in a dry place, in tightly-sealed containers between 25-30°C and protect from direct sunlight. Under optimal conditions, the medium has a shelf life of 4 years. When the container is opened for the first time, note the time and date on the label space provided on the container. After the desired amount of medium has been taken out replace the cap tightly to protect from hydration.

Product Deterioration: Do not use if they show evidence of microbial contamination, discoloration, drying or any other signs of deterioration.

DISPOSAL

Afteruse,prepared plates, specimen/sample containers and other contaminated materials must be sterilized before discarding.

REFERENCES

1. American Public Health Association, Standard Methods for the Examination of Dairy Products, 1978, 14th Ed., Washington D.C.
2. Baird R.B., Eaton A.D., and Rice E.W., (Eds.), 2015, Standard Methods for the Examination of Water and Wastewater, 23rd ed., APHA, Washington, D.C.
3. Isenberg, H.D. Clinical Microbiology Procedures Handbook. 2nd Edition.
4. Jorgensen, J.H., Pfaller, M.A., Carroll, K.C., Funke, G., Landry, M.L., Richter, S.S and Warnock., D.W. (2015) Manual of Clinical Microbiology, 11th Edition. Vol. 1.
5. Lapage S., Shelton J. and Mitchell T., 1970, Methods in Microbiology', Norris J. and Ribbons D., (Eds.), Vol. 3A, Academic Press, London.
6. MacFaddin J. F., 2000, Biochemical Tests for Identification of Medical Bacteria, 3rd Ed., Lippincott, Williams and Wilkins, Baltimore.
7. Salfinger Y., and Tortorello M.L. Fifth (Ed.), 2015, Compendium of Methods for the Microbiological Examination of Foods, American Public Health Association, Washington, D.C.



 GMP Good Manufacturing Practices Certified	 IVD For In Vitro Diagnostic Use	 QTY. Quantity	 LOT/ B. NO. Lot / Batch Number	 REF Catalogue Number	 Manufacturer
 Temperature Unit	 EC REP Authorized Representative MedNet GmbH Buckhorn 10, 48163 Münster, Germany	 European Conformity	 QR Code	 Consults Instructions for Use	 Best Before

NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

*For LabUse Only

