CM 21484 – RAPID-SENSITIVITY TEST AGAR

INTENDED USE
Forantimicrobialsusceptibility test.
PRODUCTSUMMARY ANDEXPLANATION

Thegoal ofanantimicrobial susceptibility testis to predict through an in vitro assessment the likelihood of successfully treating aninfection with a particular antimicrobial agent. There are several continual or novel methods for performing antibacterialsusceptibility testing. These include the disk diffusion test, broth microdilution, agar gradient and rapid automatedinstrument methods. Rapid-Sensitivity Test Agar, which is used for antimicrobial susceptibility tests, is a semi-defined mediumin which the mineral contents have been stabilized to give reproducible results. The thiamine and thymidinecontent is very low thus making it most suitable for testing antimicrobial activity of sulphonamides. However, some mutantstrains which are totally dependent on thiamine and thymidine for their growth, will not grow on Rapid-SensitivityTestAgar, due to very low levels of these compounds in the media as they are the naturally occurring antagonistoftrimethoprim.

COMPOSITION

Ingredients	Gms / Ltr
Casein enzymic hydrolysate	11.000
Peptic digest of animal tissue	3.000
Dextrose	2.000
Sodium chloride	3.000
Starch, soluble	1.000
Disodium phosphate	2.000
Sodium acetate	1.000
Magnesium glycerophosphate	0.200
Calcium gluconate	0.100
Cobaltous sulphate	0.001
Cupric sulphate	0.001
Zinc sulphate	0.001
Ferrous sulphate	0.001
Manganous chloride	0.002
Menadione	0.001
Cyanocobalamin	0.001
L-Cysteine hydrochloride	0.020
L-Tryptophan	0.020

Pyridoxine hydrochloride	0.003
Calcium pantothenate	0.003
Nicotinamide	0.003
Biotin	0.0003
Thiamine hydrochloride	0.00004
Adenine	0.010
Guanine	0.010
Xanthine	0.010
Uracil	0.010
Agar	8.000

PRINCIPLE

Thismedium consists of Casein enzymic hydrolysate, peptic digest of animal tissue, dextrose, and vitamins provides nitrogen, carbon compounds and other essential growth nutrients.

INSTRUCTION FOR USE

Dissolve 31.4gramsin 1000 ml purified/distilledwater.
Heat to boiling to dissolve the medium completely.
Sterilize by autoclaving at 15 psi pressure(121°C)for15minutes.
Mix well and pour into sterile petri plates.

QUALITY CONTROL SPECIFICATIONS

Appearance of Powder	: Cream to yellow homogeneous free flowing powder.
Appearance of prepared medium	: Basal medium: Light amber coloured, clear to slightly opalescent gel. After addition of 5%v/v laked blood : Red to chocolate coloured, opaque gel forms in Petri plates.
pH (at 25°C)	: 7.4 ± 0.2

INTERPRETATION

Culturalcharacteristics observed after incubation.

Microorganism	ATCC	Inoculum (CFU/ml)	Growth	Recovery	Incubation Temperature	Incubation Period
Bacillus subtilis	6633	50-100	Good-luxuriant	>=50%	35-37°C	18-24 Hours
Bacteroides vulgatus	8482	50-100	Good-luxuriant	>=50%	35-37°C	18-24 Hours

Enterococcus faecalis	29212	50-100	Good-luxuriant	>=50%	35-37°C	18-24 Hours
Salmonella Typhimurium	14028	50-100	Good-luxuriant	>=50%	35-37°C	18-24 Hours
Staphylococcus aureus	25923	50-100	Good-luxuriant	>=50%	35-37°C	18-24 Hours

PACKAGING:

Inpacksizeof100 gm and 500 gm bottles.

STORAGE

Dehydrated powder, hygroscopic in nature, store in a dry place, in tightly-sealed containers between 25-30°C and protect from direct sunlight. Under optimal conditions, the medium has a shelf life of 4 years. When the container is opened for the first time, note the time and date on the label space provided on the container. After the desired amount of medium has been taken out replace the cap tightly to protect from hydration.

Product Deterioration: Do not use if they show evidence of microbial contamination, discoloration, drying or any other signs of deterioration.

DISPOSAL

Afteruse,prepared plates, specimen/sample containers and other contaminated materials must be sterilized before discarding.

REFERENCES

1.Murray P. R., Baron J. H., Pfaller M. A., Jorgensen J. H. and Yolken R. H., (Ed.). 2003, Manual of Clinical Microbiology, 8th Ed.,American Society for Microbiology, Washington, D.C.
2. Tanner E. I. and Bullin C. H., 1974, J. Clin. Path., 27:565.
3. Thomas M. and Bond L., 1973, Med. Lab. Technol., 30:277.
4. Barker J., Healing D., and Hutchinson J. G. P., 1972, J. Clin. Path., 25:1086
5. Ericsson H. M. and Sherris J. C., 1971, Acta. Pathol. Microbiol Scand Suppl., 217:1.



NOTE: Please consult the Material Safety Data Sheet for information regarding hazards and safe handling Practices.

*For LabUse Only